

Fiber-optic pressure instrument transducers

N.D. Koshevoy, d.tech.sc., National Aerospace University
named after N.E. Zhukovskiy “KhAI”
A.V. Zabolotniy, d.tech.sc., National Aerospace University
named after N.E. Zhukovskiy “KhAI”
E.M. Kostenko, d.tech.sc., Poltavska derjavna agrarna akademiya
I.I. Koshevaya, postgraduate student, National Aerospace University
named after N.E. Zhukovskiy “KhAI”
T.G. Rozhnova, Kharkiv National University of Radioelectronics
V.V. Muratov, postgraduate student, National Aerospace University
named after N.E. Zhukovskiy “KhAI”

Kharkiv, 2019

Fiber-optic instrument transducers with elastic sensors




Fig. 1 – Fiber-optic pressure instrument transducer design

- 1 - membrane;
- 2 - case;
- 3 - transmitting optical fibers;
- 4 - central receiving optical fiber.
- 5 - photodetector;
- 6 - LEDs.




Fig. 2 – Electrical functional circuit of the instrument transducer

- 5 - photodetector;
- 6 - LEDs;
- 7 - ADC converter;
- 8 - microchip;
- 9 - digital temperature sensor;
- 10 - digital numerical indicator;
- 11 - LED driver.

Fiber optic pressure sensor using a mathematical model of the measurement process




Fig. 3 – Fiber-optic pressure instrument transducer construction

- 1 - membrane sensor;
- 2 - case;
- 3 - transmitting optical fibers;
- 4 - receiving optical fiber;
- 5 - photodetector;
- 6 - LEDs.




Fig. 4 – Electrical functional circuit of the instrument transducer

- 5 - photodetector;
- 6 - LEDs;
- 7 - ADC converter;
- 8 - microcontroller;
- 9 - temperature sensor;
- 10 - indicator;
- 11 - LED driver.

Fiber optic pressure sensor with automatic compensation for temperature error




Fig. 5 – Fiber-optic pressure instrument transducer construction

- 1 - membrane;
- 2 - case;
- 3 - transmitting optical fibers;
- 4 - receiving fiber;
- 5 - photodetector;
- 6 - LEDs.




Fig. 6 – Electrical functional circuit of the instrument transducer

- 5 - photodetector;
- 6 - LEDs;
- 7 - ADC converter;
- 8 - microcontroller;
- 9 - temperature sensor;
- 10 - digital indicator;
- 11 - LED driver;
- 12 - piezoelectric element;
- 13 - ADC converter.

Fiber optic pressure sensor with integrated control




Fig. 7 – Fiber-optic pressure instrument transducer construction

- 1 - sensor;
- 2 - case;
- 3 - transmitting optical fibers;
- 4 - receiving fiber;
- 5 - photodetector;
- 6 - LEDs.




Fig. 8 – Electrical functional circuit of the instrument transducer

- 5 - photodetector;
- 6 - LEDs;
- 7 - ADC converter;
- 8 - microcontroller;
- 9 - temperature sensor;
- 10 - digital indicator;
- 11 - LED driver;
- 12 - amplifier.

Fiber optic pressure sensor with dynamically adjustable range




Fig. 9 – Fiber-optic pressure instrument transducer:
a – functional circuit; b – electric contacts of
isolated pads connection circuit;
1- emission source, 2, 4 – optical fibers, 3 – sensor,
5 – receiver, 6 – data processing block, 7 – prism
of complete internal reflection,
8 – isolated pads, 9 – reflecting membrane, 10 –
switch block

Fiber optic pressure sensor with dynamically adjustable range and integrated control




Fig. 10– Fiber-optic instrument transducer with embedded control:
 a – functional circuit,
 b – electrical contacts of isolated pads connection circuit

Fiber optic pressure sensors for measuring the weight of moving objects




Fig. 11– Fiber-optic pressure instrument transducer's sensor design




Fig. 12 – Simplified sensor's design of a fiber-optic pressure instrument transducer

Conclusions

Fiber-optic pressure instrument transducers with elastic sensors, dynamically tuned measurement range, to measure weight of mobile objects were proposed.

Functional circuits of developed fiber-optic instrument transducers and the description of their work are given. Fiber-optic instrument transducers are protected with utility model patents of Ukraine.

Developed fiber-optic pressure instrument transducers can be widely applied in information-measuring systems and in control systems for different objects.