

CLINICAL ORAL IMPLANTS RESEARCH

*Ian Yip
Li Ma
Nikos Mattheos
Michel Dard
Niklaus P. Lang*

Defect healing with various bone substitutes

- osteoinductive
- osteoconductive
- biocompatible
- totally replaced by bone
- appropriate resorption time in relation to bone formation
- maintain graft volume stability
- satisfactory mechanical properties
- no risk of disease transmission